DEPARTMENT OF HEALTH AND HUMAN SERVICES

Public Health Service

Food and Drug Administration Nashville District Office 297 Plus Park Blvd. Nashville, TN 37217

January 21, 2000

CERTIFIED MAIL - RETURN RECEIPT REQUESTED

WARNING LETTER-00-NSV-06

July 4/00

FACILITY ID# 144543

Steve Schultz, Executive V.P. Health South Foundation University of Alabama at Birmingham The Kirklin Clinic 2000 Sixth Ave. South Birmingham, AL 35233

Dear Mr. Schultz:

Your facility was inspected on January 13, 2000 by a representative of the State of Alabama on contract to the Food and Drug Administration. This inspection revealed that your facility failed to comply with the Quality Standards for Mammography (Standards) as specified in Title 21, Code of Federal Regulations (CFR), Part 900.12, as follows:

Level 1

Phantom QC records were missing for 4 weeks for unit 1: Room Two.

Level 2

Corrective action for a failing image score (before further exams) was not documented for Unit 1, Room Two.

Corrective action for a failing image score (before further exams) was not documented for Unit 2, Room Three.

Corrective action for a failing image score (before further exams) was not documented for Unit 3, Room Six.

Corrective action for a failing image score (before further exams) was not documented for Unit 5, Room Four.

Phantom QC records were missing for at least two weeks but less than four weeks for unit 2, Room Three.

UAB - The Kirklin Clinic

Phantom QC records were missing for at least two weeks but less than four weeks for unit 3,

Phantom QC records were missing for at least two weeks but less than four weeks for unit 5, Room Four.

Responsibility for the quality assurance program and for each of its elements shall be assigned to individuals who are qualified for their assignments and who shall be allowed adequate time to perform these duties.

The Quality Control Technologist is not a qualified Mammography Technologist. The Quality Control Technologist has not been properly trained in all aspects of Mammography Q.C. testing.

These specific deficiencies appear on the Post Inspection Report, which was given to your facility at the close of your inspection. These deficiencies are symptomatic of serious problems that could compromise the quality of mammography at your facility and potentially overexpose both patients and employees involved with mammography.

It is your responsibility to ensure adherence to each requirement of the Mammography Quality Standards Act of 1992 (MQSA) and FDA's regulations. You are responsible for investigating and determining the cause of these deficiencies that the inspection identifies and to promptly initiate permanent corrective actions.

If you fail to properly address these deficiencies, FDA may, without further notice, initiate regulatory action. Under MQSA, FDA may:

- impose civil money penalties on a facility of up to \$10,000 for each failure to substantially comply with, or each day of failure to comply with the Standards.
- suspend or revoke a facility's FDA certificate for failure to comply with the Standards.
- seek an injunction in federal court to prohibit any mammography activity that constitutes a serious risk to human health.

Within 15 working days after receiving this letter, you should notify FDA in writing of each step your facility is taking to prevent the recurrence of similar violations.

If your facility is unable to complete the corrective actions within 15 working days, you should state the reason for the delay and the time within which the corrections will be completed.

ч. _Х

UAB - The Kirklin Clinic

Your reply should be directed to Joseph E. Hayes, Compliance Officer, U.S. Food and Drug Administration, 297 Plus Park Boulevard, Nashville, Tennessee 37217, telephone 615/781-5389, extension 125. Any questions in regard to this letter or how to ensure you are meeting MQSA standards, please call Karen Smallwood, Radiation Specialist, at 615/781-5380, extension 144.

Sincerely,

Howard E. Lewis,

Acting Director

New Orleans District Office

HEL/krs

cc: State of Alabama